

MASCARILLA HIGIÉNICA - de un solo uso: GB/T32610-2016
USO PREVISTO

Esta mascarilla higiénica desechable es de uso general para la población civil sana, eficaz para obstruir y evitar la propagación de gotas, polvo y gérmenes de la gripe. Para proteger a las personas de los agentes infecciosos, contra salpicaduras o estornudos de líquidos y partículas contaminadas. Uso individual, no estéril.

Test estándar aplicable según la normativa GB/T32610-2016

Características	Métodos de prueba	Limites Aceptables
Eficiencia filtrando (medio soldado)	GB/T32610-2016	Requerimiento $\geq 90\%$ - Resultado del test: 94,6%
Resistencia Inspiración	GB/T32610-2016	Requerimiento ≤ 175 - resultado del test: 95
Resistencia Expiración	GB/T32610-2016	Requerimiento ≤ 145 - resultado del test: 78

Características		Especificaciones	
Materiales	Cuerpo de la mascarilla (3 capas)	Capa Exterior	Tejido No-tejido, 25 gs/m ²
		Filtro	N95 Polipropileno Fundido, 25 gs/m ²
		Filtro interior	Tejido No tejido, 25 gs/m ²
	Pieza de la nariz	Pieza metálica recubierta de plástico	
	Tiras para las orejas	Algodón y Licra	
Descripción	Mascarilla con pliegues rectangulares, con una pieza ajustable para la nariz con dos tiras, una a cada lado para poder ajustarla al rostro.		
Tipo de asas	Tiras confeccionadas en tejido plano para sujetar a las orejas		
Medidas	Cuerpo de la mascarilla	Longitud	175 mm
		Ancho	95 mm
	Longitud de las asas	Normal	190 mm (± 5 mm)
	Profundidad del pliegue		13 mm
	Longitud de la pieza de la nariz		95 mm

Información adicional

Condiciones de almacenaje	País de origen
Mantener en un lugar fresco, limpio y seco. Evitar temperaturas altas (más 40°C o 104°F)	China





8 437013 532143
código de barras
PACK 10 unidades



Caja Master



10 mascarillas por packs




8 437013 532150
código de barras estuche/50 mascarillas

Apéndice 1 lista de códigos de artículo y atributos

Código Interno	Descripción	Información envasado	Color	Atributo específico	Tratamiento	Medidas caja cm	Peso neto/bruto
258HIGI-ENICA0	Mascarilla Facial Desechable	10 mascarillas/pack 5 packs/estuche 40 estuches/caja master 2.000 mascarillas/caja	Azul	Asas longitud regular	No estéril	50 x 40 x 35	7.5/8.5 KGS

FOTOS



9.5cm

Product Certificate 产品合格证

Product Name: Disposable face mask (non-medical)

产品: 非医用一次性防护口罩

Specification: 17.5cm x 9.5cm 规格: 17.5cm x 9.5cm

Major ingredient: 70% non woven fabric and 30% melt-blown polypropylene

主要成分: 70%无纺布和30%熔喷布

Executive standard: GB/T32610-2016

执行标准: GB/T32610-2016

Production date: 5th MAY-2020

生产日期: 2020年05月05日

Period of validity: 2 years 有效期: 二年

Batch code: HJ/20200505 批次号: HJ/20200505

Inspector: Du Peng 检验员: 杜鹏

Manufacturer: WENZHOU HENGJI PACKING CO. LTD

生产厂商: 温州恒基包装有限公司

The product has been tested and proved to be qualified in conformity with standard and is approved for sales and use.

本产品经检验符合规定的质量标准要求, 准予出厂。



5.4cm



180011112242



(2018) 国认监认字 (244) 号



中国认可
国际互认
检测
TESTING
CNAS L6780

检 验 报 告

TEST REPORT



Scan is following us

Scan checks for authenticity



* W 2 0 2 0 0 5 6 9 8 E *

浙江省轻工业产品质量检验研究院

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

Test report

Number:W202005698E

page1/3

Name of Customer	WENZHOU HENGJI PACKING CO., LTD	Address	DANPU INDUSTRY ZONE. , XIAOJIANG , PINGYANG , WENZHOU , ZHEJIANG PROVINCE, CHINA
Manufacturer	---	Address	---
Sample information	Name of sample: face mask Characteristics of sample: blue Trademark of sample: --- Specification/model: --- Level: --- Category of safety specification: --- Art. No. : --- -----		
Above-mentioned information by Customer-supplied			
The sent way of sample	Courier	Sample quantity	One package
Receiving Date of Sample	2020/04/12	Test Category	Entrusted inspection
Date of Testing	2020-04-13- 2020-04-13		
RatingRequirements	GB/T 32610-2016		
Test Summary: See the attached page for the results. <div style="text-align: right;"> Test Seal Date of Approval: 2020-04-14 </div>			
Remarks	1.The filter efficiency grade is not marked on the sample mark, and it is judged according to the lowest standard requirement III grade. 2. The sample is not pretreated and the result is for reference.		

Approved by: 曹 丽 勤

Test report

Number:W202005698E

page2/3

ITEM	STANDARD (D GRADE)	RESULT	RATING
1.FILTER EFFICIENCY(SALT MEDIUM) (GB/T32610-2016)(%)			
-	≥ 90	94.6	PASS
2. RESPIRATORY RESISTANCE (GB/T32610-2016)(Pa)			
-INSPIRATORY RESISTANCE	≤ 175	95	PASS
-EXPIRATORY RESISTANCE	≤ 145	78	PASS



Test report

Number:W202005698E

page3/3

Picture(s) of sample



—End of report—



DECLARATION

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
2. The test report will be deemed invalid without signatures of the inspector/reviewer and authorized personnel, and the red special inspection stamp of our organization.
3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
4. The test results shown in this report is only valid for the tested sample.
5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

浙江省轻工业品质量检验研究院及附设的检验中心 The Affiliated Inspection Centers

浙江省轻工业品质量检验研究院

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center (Zhejiang)

地址：浙江省杭州市江干区下沙路 300 号 6 号楼

Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang

联系电话：0571-85122669

Telephone: 0571-85122669

E-Mail: 1137907994@qq.com

国家家具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Furniture (Zhejiang)

浙江省室内安全及家具产品质量检验中心

Zhejiang Center of Quality Test for Indoor Safety and Furniture Products

地址：浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 4 号楼

Address: Building 4 LiangZhu University Science and Technology Park, No.1,
JingYi Rd, Yuhang District, Hangzhou, Zhejiang

联系电话：0571-89009556

Telephone: 0571-89009556

E-Mail: 2047699564@qq.com

浙江省轻工及五金产品质量检验中心

Zhejiang Center of Quality Test for Light Industry and Hardware Products

浙江省体育用品质量检验中心

Zhejiang Center of Quality Test for Sports Products

地址：浙江省杭州市西湖区天目山路 222 号 3 号楼

浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 3 号楼

Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang

Building 3 LiangZhu University Science and Technology Park, No.1,
JingYi Rd., Hangzhou, Zhejiang

联系电话：0571-89001107

Telephone: 0571-89001107

E-Mail: wujtest@126.com

国家锁具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Lock (Zhejiang)

浙江省锁具产品质量检验中心

Zhejiang Center of Quality Test for Lock Products

地址：浙江省杭州市塘苗路 24 号

Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang

联系电话：0571-85027738

Telephone: 0571-85027738

E-Mail: Locktest@126.com

网上业务受理/报告查询: <http://www.zjttj.cn:807>

Online Business-Reception/Report Inquires: <http://www.zjttj.cn:807>

投诉电话：0571-85023552

Complaint Tel: 0571-85023552